



## Clinical implications

Factor XIII, also called fibrin stabilizing factor, is a proenzyme circulating in plasma. Its active form stabilizes the fibrin network by catalytic formation of covalent bonds between fibrin molecules and other proteins (alpha-2-antiplasmin, fibronectin, collagen), thereby protecting the coagulum against fibrinolysis. FXIIIa plays an important role not only in hemostasis but also during wound healing and in maintaining pregnancy. The presence of the V34L has a protective effect against myocardial infarction and venous thrombosis. On the other hand, it is also stated that the presence of this

polymorphism is associated with an increased risk of recurrent pregnancy loss in the first trimester, which was also found in connection with the PAI-1 (4G/5G) mutation.

## Principle of detection

The kit is intended for detection of mutation V34L in fibrin stabilizing factor XIII in human genomic DNA. Detection is based on **real-time polymerase chain reaction (qPCR) using fluorescently labelled probes (allelic discrimination)**.

## Available products

Cat. No.	Product	rxn
3209-025	gb HEMO FXIII (V34L)	25
3209	gb HEMO FXIII (V34L)	100

1 kit contains reagents to provide 25 or 100 PCR reactions (20 µl volume of each reaction).


## Parameters of the diagnostic kit

- *in vitro* diagnostics
- CE IVD marked
- ready-to-use assay
- sample concentration 1-100 ng/µl
- positive and negative controls included
- FAM and HEX channels detection
- identical amplification profile as gb HEMO, gb GENETIC, gb PHARM kits

## Content of the diagnostic kit

* Component	Conc.	Purpose
Assay qPCR F13 (V34L)	1.25×	Detection assay
Deionized Water		Negative Control
Standard WT F13 (V34L)	10 <sup>4</sup> cop/µl	Positive Control
Standard MUT F13 (V34L)	10 <sup>4</sup> cop/µl	Positive Control
Standard HET F13 (V34L)	10 <sup>4</sup> cop/µl	Positive Control

\* Lid colour



## Validated for cyclers

- Rotor-Gene 3000/6000/Q (Corbett Research, Qiagen)
- iCycler iQ5/CFX96/CFX96 Touch (Bio-Rad)
- ABI 7500/7500 Fast/7900HT (Applied Biosystems)
- SmartCycler (Cepheid)
- MIC (Bio Molecular Systems)
- AriaMx (Agilent Technologies)
- Light Cycler 480/Cobas z480 (Roche Diagnostics)

Fig. 1 – Detection of FXIII (V34L) standards on CFX96 device; blue line – wild type; red line – mutant; green line – heterozygote; smooth line – FAM channel; dotted line – HEX channel